Yursat Dış Tic. Mühendislik
Yeni Karaman Mah. Sanayi Cad. 4. Kantar Sok. Asya Plaza Kat:5 No:505
Osmangazi/BURSA
info@yursat.com.tr
www.yursat.com.tr

Alpha Moisture System

Garanti: 12 ay

Bu ürüne ait Türkiye içi kapı teslim fiyat ve teslim süresi içeren teklifimizi almak için **info@yursat.com.tr** e-posta adresine ya da www.yursat.com.tr/iletisim adresindeki iletişim formu üzerinden bizlere ulaşabilirsiniz.

Alpha Moisture System Markası, tedarik süresi için lütfen bizimle iletişime geçiniz.

Firmamız Alpha Moisture System Türkiye Distribütörü veya temsilcisi değildir. Firmamız sipariş durumunda, belirtilen ürünlerde sadece Orjinal ve yeni ürün teklifi sunmaktadır. Bu sitede gösterilen Özel marka adları ve ticari markalar ilgili sahiplerinin mülkiyetindedir, talep durumunda kaldırılmaktadır.

Ürün	Açıklama
MT-706260	Sensor Type : Ultra high capacitance AlOx transmitter Measuring Range : -90°C to 0°C dewpoint Accuracy : ±2°C dewpoint Operating Temperature : -20°C to +60°C Operating Pressure : 0.01 to 350 bara Operating Temperature : -20 to +60°C. AutoCal : Field calibration/span check operated via 2 buttons on transmitter body. Locked under normal operation to avoid tamper. Supply Voltage : 12 to 28 VDC Output Signal : Linear 4-20mA Included : Operating Instruction Manual. 2 metres of connecting cable. Certificate of Test and Calibration (traceable to National/International Humidity Standards) The model AMT Dewpoint Transmitter for safe area use is a two-wire transmitter, used for continuous measurement of moisture in dry process gases, which will output a linear 4-20mA signal corresponding to the dewpoint temperature. The Ultra High Capacitance sensing element is long lasting and offers excellent sensitivity, repeatability, and response speed. Each unit is supplied with a Certificate of Calibration, traceable to International Humidity Standards, validating accuracy better than ±2°C dewpoint. The AMT also has the unique AutoCal feature which allows users of the Alpha Moisture Systems instruments to ensure optimum accuracy between laboratory calibrations by checking the span of the instrument and correcting if required.
4314 KRC	
Sonda Di Rugiada	